



RESTRICTED - COMMERCIAL

Radiation Survey of Lonestar Unit  
for  
Owlstone Limited

May 2008



**Report Title:** Radiation Survey of Lonestar Unit



**Customer Contact:** Max Allsworth  
NPI Manager

**Address:** Owlstone Ltd  
St John's Innovation Centre  
Cowley Road  
Cambridge  
CB4 0WS  
United Kingdom

**Telephone:** + 44 1223 422415

**Date of Survey:** 21 April 2008  
**Date of Report:** 30 May 2008  
**Report Ref.:** AHP/RPA/OWL/REP/o8/o1

**Please Contact:** Paddy Copeland  
**Telephone:** +44 1235 838622  
**Fax:** +44 1235 838623  
**Mobile:** +44 7917 804466  
**Email:** paddy.copeland@aurorahp.co.uk

	<b>Name</b>	<b>Position</b>	<b>Date</b>	<b>Signature</b>
<b>Author</b>	Paddy Copeland	Radiation Protection Adviser	30 May 2008	
<b>Reviewer</b>	Glenn Hardcastle	Radiation Protection Adviser	2 June 2008	

**CONTENTS**

**1. INTRODUCTION .....4**

**2. RADIATION SURVEY.....4**

**2.1 Survey Methodology .....4**

**2.2 Survey Results .....4**

**3. CONCLUSIONS .....6**

**APPENDIX 1 CALIBRATION CERTIFICATES ..... 7**

## 1. INTRODUCTION

This report relates to a radiation survey of the external surface of one of Owlstone Limited's Lonestar units (serial number 00018) carried out by Aurora Health Physics Services Ltd. Owlstone is planning to sell Lonestar units commercially and asked Aurora to carry out an independent survey prior to bringing the unit to market.

The Lonestar unit contains a single 555 MBq nickel-63 source. Nickel-63 emits beta radiation with a maximum energy of 66 keV.

## 2. RADIATION SURVEY

### 2.1 Survey Methodology

The radiation survey undertaken comprised of direct monitoring, in contact with, and at a number of fixed distances from all external surfaces of the Lonestar unit using the following instruments:

**Table 1: Instruments used.**


Instrument	Serial No(s)	Calibration Date	Background (on site)
Mini Instruments 900 D	AHP0041 (4368)	29/05/07	<0.1 µSv/h
Mini Instruments 900 EP15	AHP0077 (2269)	16/11/07	1 - 2 counts per second (cps)
Thermo Electron Electra/BP19DD	AHP0101 (3803/386)	24/12/07	10 - 15 cps
Automess 6150 AD2	AHP0061 (73796)	3/07/07	0.05 µSv/h

Instrument calibration certificates are appended to this report (see Appendix 1).

### 2.2 Survey Results

The direct radiation monitoring of all external surfaces of the Lonestar unit detected localised elevated radiation readings over a small area on the rear of the unit. The maximum observed contact dose rate was 0.8 µSv/h. Radiation readings over all other surfaces of the unit were at background levels. Radiation readings at 5cm from the unit were all at background levels

(less than 0.1  $\mu\text{Sv/h}$ ). Figure 1 below shows the measured dose-rate profile around the unit. Dose rates in the vertical plane follow the same profile.



**Figure 1. Dose rate profile around Lonestar Unit.**

### 3. CONCLUSIONS

A dose rate survey has been carried out around a Lonestar Unit (serial number 00018). The maximum measured dose rate in contact with the unit was 0.8  $\mu\text{Sv/h}$ . Dose rates at 5cm were all at background levels (less than 0.1  $\mu\text{Sv/h}$ ).

**APPENDIX 1 CALIBRATION CERTIFICATES**



The world leader  
in serving science

<b>CERTIFICATE OF CALIBRATION</b>			
<b>ISSUED BY:</b>	<b>Thermo Fisher Scientific RM&amp;SI Beenham Radiation Calibration Laboratory</b>		
<b>DATE OF ISSUE:</b>	29 May 2007	<b>CERTIFICATE No:</b>	9125

**APPROVED SIGNATORY:** **J. Lowe**  
 This certificate details the calibration results obtained for the instrument named below when calibrated in conjunction with calibration procedure RCS 028

The procedures used for the measurements are consistent with GPG14 (1999).  
 All measurements are checked by a designated 'Qualified Person' as defined by regulation 19(3) of the Ionising Radiation Regulations 1999.

**Instrument Type:** Mini 900 D AMB  
**Instrument Manufacturer:** Thermo Fisher Scientific  
**Instrument Serial Number:** 4368  
**Radiation Type Measured:** Gamma Dose Rate (Sv/h)  
**Customer:** Thermo Fisher Scientific

**TRACEABILITY**  
 All measurements were performed with a gamma source system which has been calibrated with cross referenced ionisation chambers and electrometers in accordance to BS 5869 (1980), ISO 4037 (1979), implying traceability to National Standards.

**UNCERTAINTIES**  
**System Dose**  
<sup>137</sup>Cs and <sup>60</sup>Co: For air kerma rates above 1 μGy.h<sup>-1</sup>, the uncertainty in the system dose rate is ± 3%.  
<sup>241</sup>Am: For air kerma rates above 20 μGy.h<sup>-1</sup>, the uncertainty in the system dose rate is ± 4%.  
**Indicated Dose Rate**  
 All indicated dose rates quoted, except the background response, have an uncertainty of ±10%.

**INTRINSIC ERROR**  
 The intrinsic error is calculated by  $\frac{\text{indicated dose rate} - \text{system dose rate}}{\text{system dose rate}} \times 100\%$

**NOTE:** This calibration is not automatically valid for conditions of use different to those of calibration as given upon this certificate.

The uncertainties are for a confidence probability of not less than 95%. **P.T.O.**

**CERTIFICATE OF CALIBRATION**

DATE OF ISSUE: 29 May 2007 CERTIFICATE No: 9125

**Measurement Results**

The Probe was mounted horizontally, facing the gamma radiation beam.

Date of Measurements: 29 May 2007  
 Instrument Type: Mini 900 D AMB  
 Instrument Serial Number: 4368  
 Test Type: Initial  
 Temperature (°C): 21°C  
 Background Response ( $\mu\text{Sv}\cdot\text{h}^{-1}$ ):  $<0.3\mu\text{Sv/h}$   
 Pressure (mbar): 1004mbar

**Linearity Response  $^{137}\text{Cs}$**

Source Number	System Dose Rate	Indicated Dose Rate	Intrinsic Error
4291GR	300 $\mu\text{Sv/h}$	295 $\mu\text{Sv/h}$	-1.67%
7581GR	20 $\mu\text{Sv/h}$	20 $\mu\text{Sv/h}$	0.00%
7581GR	2 $\mu\text{Sv/h}$	2.2 $\mu\text{Sv/h}$	10.00%

**Overload Response**


Off scale for 10 secs: Satisfactory

**Energy and Polar Response  $^{241}\text{Am}$**

Source Number	Orientation	System Dose Rate	Indicated Dose Rate	Intrinsic Error
1498LX	0°	100 $\mu\text{Sv/h}$	90 $\mu\text{Sv/h}$	-10.00%
1498LX	-90°	100 $\mu\text{Sv/h}$	62 $\mu\text{Sv/h}$	-38.00%

Remarks: [1] All system dose and dose rates are ambient dose equivalent rates, H\*(10).

Measurements Performed By:


 J.W.V. Coles

The uncertainties are for a confidence probability of not less than 95%.



**CALIBRATION CERTIFICATE**

Radiation Metrology Group, Radiation Protection Division  
Health Protection Agency, Chilton, Didcot, Oxfordshire, OX11 0RQ  
E-mail: rad.met@hpa.org.uk Web: www.hpa.org.uk/radiation  
Tel: +44(0)1235 831 600 Fax: +44(0)1235 822 818



**Radiological Protection Instrument  
Certificate of Calibration  
No. 49024**

**A. Name and address of user:**

Aurora Health Physics Ltd  
Harwell Innovation Centre  
Curle Avenue  
TIDCOT  
Oxfordshire  
OX11 0QG

**B. Details of instrument and use:**

Description: **Thermo 900 with type EP15 probe**  
Serial No: **2269 and -- respectively**  
Use: **Monitoring of surfaces contaminated with beta emitting nuclides.**

**C. Details of calibration:**

Date: **16 November 2007**  
Type of calibration: **Annual calibration**  
Calibration conditions: **The instrument's linearity was determined using a range of  $^{90}\text{Sr}$  +  $^{90}\text{Y}$  ISO extended area contamination plaque sources. The response to a variety of other nuclides was also measured.**  
Source to detector separation: **3 mm from the front of the detector**  
Orientation: **Radiation beam perpendicular to detector window**  
Uncertainties: **The uncertainties stated in this certificate are expressed at the 95% confidence level.**  
Traceability: **All source activities quoted in this certificate were derived from emission rate measurements using a P factor of 2. All emission rate measurements are directly traceable to national standards via UKAS accredited calibration facilities.**

**D. Results of calibration:**

**Satisfactory for measurements performed  
See over**

**E. Compliance with the Ionising Radiations Regulations 1999:**

**These results are typical of type and the instrument is suitable for the use described in section B, under the terms of the Regulations and the associated Approved Code of Practice.**

Calibration performed by: Miss C S Beswick Signature: *C S Beswick*  
Approved by (Qualified Person): Mr T J Daniels Signature: *T J Daniels*  
Date of issue: 13 November 2007

**CALIBRATION**

**Certificate of calibration No. 49024**

**Results of calibration:**

**F. Physical inspection:**      **Battery test:**                      **Satisfactory**  
    **Condition of Instrument:**      **Satisfactory**  
    **Audio:**                                      **Satisfactory**

**G. Linearity:**

Source Activity ( <sup>90</sup> Sr + <sup>90</sup> Y)	Instrument Range	Instrument Reading (inc BG)	Response Factor	Estimated Uncertainty
Background	0 - 2000 Counts s <sup>-1</sup>	0	-----	±50%
3.2 Bq cm <sup>-2</sup>	0 - 2000 Counts s <sup>-1</sup>	14	4.4	±20%
16.4 Bq cm <sup>-2</sup>	0 - 2000 Counts s <sup>-1</sup>	70	4.3	±20%
54.7 Bq cm <sup>-2</sup>	0 - 2000 Counts s <sup>-1</sup>	240	4.4	±20%

**H. Other radiation responses:**

Nuclide	Instrument Range	Instrument Reading (inc BG)	Response Factor	Estimated Uncertainty
Background	0 - 2000 Counts s <sup>-1</sup>	0	-----	±50%
<sup>36</sup> Cl	0 - 2000 Counts s <sup>-1</sup>	140	3.8	±20%
<sup>60</sup> Co	0 - 2000 Counts s <sup>-1</sup>	55	2.4	±20%
<sup>147</sup> Pm	0 - 2000 Counts s <sup>-1</sup>	13	1.6	±20%
<sup>14</sup> C	0 - 2000 Counts s <sup>-1</sup>	240	1.7	±20%
<sup>241</sup> Am	0 - 2000 Counts s <sup>-1</sup>	50	2.3	±20%

Response factor is defined as the factor by which the indication of the instrument has to be divided by to give the units as defined below.

$$\text{Response factor} = \frac{\text{Indicated reading} - \text{BG, counts s}^{-1}}{\text{Estimated source activity} - \text{BG, Bq cm}^{-2}}$$

It is estimated that for monitoring of surfaces uniformly contaminated with:

- <sup>32</sup>P      4.4 cps above background = 1 Bqcm<sup>-2</sup>
- <sup>33</sup>P      1.6 cps above background = 1 Bqcm<sup>-2</sup>
- <sup>36</sup>S      1.7 cps above background = 1 Bqcm<sup>-2</sup>
- <sup>48</sup>Ca     1.6 cps above background = 1 Bqcm<sup>-2</sup>
- <sup>204</sup>Tl     3.8 cps above background = 1 Bqcm<sup>-2</sup>

**END OF DATA**

CALIBRATION CERTIFICATE

Radiation Metrology Group, Radiation Protection Division  
Health Protection Agency, Chilton, Didcot, Oxfordshire, OX11 0RQ  
E-mail: rad.met@hpa.org.uk Web: www.hpa.org.uk/radiation  
Tel: +44(0)1235 831 600 Fax: +44(0)1235 822 818



**Radiological Protection Instrument  
Certificate of Calibration  
No. 49498**

**A. Name and address of user:**

Aurora Health Physics Ltd  
Harwell Innovation Centre  
Curie Avenue  
Didcot  
Oxfordshire  
OX11 0QG

**B. Details of instrument and use:**

Description: NE Electra 1A with type BP19DD probe  
Serial No: 3803 and 386 respectively  
Use: Monitoring of surfaces contaminated with beta emitting radionuclides.

**C. Details of calibration:**

Date: 24 December 2007  
Type of calibration: Annual calibration  
Calibration conditions: The instrument's linearity was determined using a range of <sup>90</sup>Sr + <sup>60</sup>Co ISO extended area contamination plaque sources. The response to a variety of other nuclides was also measured.  
Source to detector separation: 3 mm from the front of the detector  
Orientation: Radiation beam perpendicular to detector window  
Uncertainties: The uncertainties stated in this certificate are expressed at the 95% confidence level.  
Traceability: All source activities quoted in this certificate were derived from emission rate measurements using a P factor of 2. All emission rate measurements are directly traceable to national standards via UKAS accredited calibration facilities.

**D. Results of calibration:**

Satisfactory for measurements performed  
See over

**E. Compliance with the Ionising Radiations Regulations 1999:**

These results are typical of type and the instrument is suitable for the use described in section B, under the terms of the Regulations and the associated Approved Code of Practice.

Calibration performed by: Mr D E Payne      Signature: *DE Payne*  
Approved by (Qualified Person): Mr D R Bungay      Signature: *[Signature]*  
Date of Issue: 24 December 2007

CALIBRATION

**Certificate of calibration No. 49498****Results of calibration:**

F. Physical inspection:      Battery test:                      Satisfactory  
    Condition of Instrument:      Satisfactory  
    Audio:                                      Satisfactory  
    E.H.T.:                                      980V

**G. Linearity:**

Source Activity ( <sup>90</sup> Sr + <sup>90</sup> Y)	Instrument Range	Instrument Reading (inc BG)	Response Factor	Estimated Uncertainty
Background	Digital Counts s <sup>-1</sup>	6.83	---	±50%
3.2 Bq cm <sup>-2</sup>	Digital Counts s <sup>-1</sup>	89	25.90	±20%
16.5 Bq cm <sup>-2</sup>	Digital Counts s <sup>-1</sup>	451	27.15	±20%
54.9 Bq cm <sup>-2</sup>	Digital Counts s <sup>-1</sup>	1540	28.12	±20%

**H. Other radiation responses:**

Nuclide	Instrument Range	Instrument Reading (inc BG)	Response Factor	Estimated Uncertainty
Background	Digital Counts s <sup>-1</sup>	6.83	---	±50%
<sup>36</sup> Cl	Digital Counts s <sup>-1</sup>	955	25.97	±20%
<sup>60</sup> Co	Digital Counts s <sup>-1</sup>	430	18.71	±20%
<sup>147</sup> Pm	Digital Counts s <sup>-1</sup>	115	13.59	±20%
<sup>14</sup> C	Digital Counts s <sup>-1</sup>	1931	13.49	±20%
<sup>241</sup> Am	Digital Counts s <sup>-1</sup>	484	22.22	±20%

Response factor is defined as the factor by which the indication of the instrument has to be divided by to give the units as defined below.

$$\text{Response factor} = \frac{\text{Indicated reading} - \text{BG, counts s}^{-1}}{\text{Estimated source activity} - \text{BG, Bq cm}^{-2}}$$

It is estimated that for monitoring of surfaces uniformly contaminated with:

<sup>32</sup> P	29.2 cps above background = 1 Bqcm <sup>-2</sup>
<sup>33</sup> P	12.6 cps above background = 1 Bqcm <sup>-2</sup>
<sup>35</sup> S	11.4 cps above background = 1 Bqcm <sup>-2</sup>
<sup>46</sup> Ca	12.6 cps above background = 1 Bqcm <sup>-2</sup>
<sup>204</sup> Tl	2.9 cps above background = 1 Bqcm <sup>-2</sup>

END OF DATA

CALIBRATION CERTIFICATE



Radiation Metrology Group  
National Radiological Protection Board, Chilton, Didcot, Oxon, OX11 0RQ  
E-mail: rad.met@nrpb.org Web: www.nrpb.org  
Tel: (01235) 831600 Fax (01235) 822818

**Radiological Protection Instrument  
Certificate of Calibration  
No. 47551**

**A. Name and address of user:**

**Aurora Health Physics Ltd**  
Harwell Innovation Centre  
Curie Avenue  
Chilton  
Oxfordshire  
OX11 0QG

**B. Details of instrument and use:**

Description: **Automess 6150 AD**  
Serial No: **73796**  
Use: **x and  $\gamma$  radiation above 60 keV**

**C. Details of calibration:**

Date: **3 July 2007**  
Type of calibration: **Annual calibration**  
Calibration conditions: **The instrument's linearity was tested over a range of ambient dose equivalent rates using  $^{137}\text{Cs}$ . The instrument's response to  $^{241}\text{Am}$  and  $^{60}\text{Co}$  was also measured and the overload and dose range performance were confirmed.**  
  
Calibration reference point: **As marked**  
Orientation: **Radiation beam normal to front face of instrument**  
Uncertainties: **The uncertainties stated in this certificate are expressed at the 95% confidence level.**  
Traceability: **All the equipment associated with the measurements performed in this certificate have calibrations directly traceable to national standards via the National Physical Laboratory or UKAS accredited calibration facilities.**

**D. Results of calibration:**

Satisfactory for measurement performed.  
See over.

**This instrument is designed to measure the quantity photon dose equivalent, H<sub>p</sub>. The calibration was performed using the quantity ambient dose equivalent and therefore the responses obtained for  $^{241}\text{Am}$  and  $^{60}\text{Co}$  are as anticipated.**

**E. Compliance with the Ionising Radiations Regulations 1999:**

These results are typical of type and the instrument is suitable for the use described in section B, under the terms of the Regulations and the associated Approved Code of Practice.

Calibration performed by: Mr D E Payne Signature:

Approved by (Qualified Person): Mr D R Bungay Signature:

Date of issue: 3 July 2007

CALIBRATION

**Certificate of calibration No. 47551**

**Results of calibration:**

**F. Physical inspection:** Instrument condition: Satisfactory  
 Battery test: Satisfactory  
 Audio: Satisfactory

**G. Instrument settings:** Alarm operation: Satisfactory

**H. Dose Rate linearity;**

Radiation Quality	Instrument Range	Instrument Reading	Response Factor	Estimated Uncertainty
Background	0-10 $\mu\text{Sv h}^{-1}$	0.04	--	$\pm 10\%$
<sup>137</sup> Cs (662 keV)	0-10 $\mu\text{Sv h}^{-1}$	2.4	1.12	$\pm 10\%$
<sup>137</sup> Cs (662 keV)	0-10 $\mu\text{Sv h}^{-1}$	5.69	0.98	$\pm 10\%$
<sup>137</sup> Cs (662 keV)	0-100 $\mu\text{Sv h}^{-1}$	18.4	1.00	$\pm 10\%$
<sup>137</sup> Cs (662 keV)	0-100 $\mu\text{Sv h}^{-1}$	62.4	1.02	$\pm 10\%$
<sup>137</sup> Cs (662 keV)	0-1000 $\mu\text{Sv h}^{-1}$	197	1.00	$\pm 10\%$
<sup>137</sup> Cs (662 keV)	0-1000 $\mu\text{Sv h}^{-1}$	572	1.01	$\pm 10\%$
<sup>137</sup> Cs (662 keV)	0-10 $\text{mSv h}^{-1}$	2.19	0.99	$\pm 10\%$
<sup>137</sup> Cs (662 keV)	0-10 $\text{mSv h}^{-1}$	6.62	0.98	$\pm 10\%$

**I. Dose Response:**

<sup>137</sup> Cs (662 keV)	$\mu\text{Sv}$	50	1.00	$\pm 10\%$
-----------------------------	----------------	----	------	------------

**J. Other nuclide responses:**

<sup>241</sup> Am (60 keV)	0-1000 $\mu\text{Sv h}^{-1}$	94.5	0.37	$\pm 10\%$
<sup>60</sup> Co (1.25 MeV)	0-1000 $\mu\text{Sv h}^{-1}$	528	1.35	$\pm 10\%$

**K. Overload:**

<sup>137</sup> Cs (662 keV)	0-10 $\text{mSv h}^{-1}$	Overload satisfactory at > 63 $\text{mSv h}^{-1}$
-----------------------------	--------------------------	---

Response factor is defined as the factor by which the indication of the instrument has to be divided by to give the units as defined below. These factors are referred where applicable to a temperature and pressure of 20°C and 1013 mbar.

$$\text{Response factor} = \frac{\text{Indicated reading} - \text{BG, } \mu\text{Sv h}^{-1}, \text{ mSv h}^{-1}, \mu\text{Sv}}{\text{True ambient dose equivalent rate} - \text{BG, } \mu\text{Sv h}^{-1}, \text{ mSv h}^{-1}, \mu\text{Sv}}$$

**END OF DATA**